



US 20180364145A1

(19) **United States**

(12) **Patent Application Publication**
Rich

(10) **Pub. No.: US 2018/0364145 A1**

(43) **Pub. Date: Dec. 20, 2018**

(54) **SYSTEMS AND USER INTERFACE FOR COLLECTING A DATA SET IN A FLOW CYTOMETER**

G01J 3/02 (2006.01)
G01J 3/36 (2006.01)
G01J 3/18 (2006.01)
G06M 11/02 (2006.01)

(71) Applicant: **Becton, Dickinson and Company**, Franklin Lakes, NJ (US)

(52) **U.S. Cl.**
CPC *G01N 15/1434* (2013.01); *G01N 15/1404* (2013.01); *G06F 19/26* (2013.01); *G06F 19/12* (2013.01); *G06F 19/10* (2013.01); *G06M 11/02* (2013.01); *G01J 3/0208* (2013.01); *G01N 15/14* (2013.01); *G01J 3/36* (2013.01); *G01J 3/18* (2013.01); *G01J 3/0229* (2013.01); *G01N 15/1429* (2013.01)

(72) Inventor: **Collin A. Rich**, Ann Arbor, MI (US)

(21) Appl. No.: **16/015,807**

(22) Filed: **Jun. 22, 2018**

Related U.S. Application Data

(60) Continuation of application No. 15/004,521, filed on Jan. 22, 2016, now Pat. No. 10,031,064, which is a division of application No. 13/820,968, filed on Apr. 4, 2013, now Pat. No. 9,280,635, filed as application No. PCT/US11/57747 on Oct. 25, 2011.


(60) Provisional application No. 61/406,255, filed on Oct. 25, 2010, provisional application No. 61/406,251, filed on Oct. 25, 2010, provisional application No. 61/406,259, filed on Oct. 25, 2010.

Publication Classification

(51) **Int. Cl.**
G01N 15/14 (2006.01)
G06F 19/26 (2011.01)
G06F 19/12 (2011.01)
G06F 19/10 (2011.01)

(57) **ABSTRACT**

Systems in a flow cytometer having an interrogation zone and illumination impinging the interrogation zone include: a lens subsystem including a collimating element that collimates light from the interrogation zone, a light dispersion element that disperses collimated light into a light spectrum, and a focusing lens that focuses the light spectrum onto an array of adjacent detection points; a detector array, including semiconductor detector devices, that collectively detects a full spectral range of input light signals, in which each detector device detects a subset spectral range of the full spectral range of light signals; and a user interface that enables a user to create a set of virtual detector channels by grouping detectors in the detector array, such that each virtual detector channel corresponds to a detector group and has a virtual detector channel range including the sum of subset spectral ranges of the detectors in the corresponding detector group.






FIGURE 1




FIGURE 2




FIGURE 3




FIGURE 4A






FIGURE 4B





S320





FIGURE 6B

FIGURE 6C

FIGURE 6D

FIGURE 6E



SYSTEMS AND USER INTERFACE FOR COLLECTING A DATA SET IN A FLOW CYTOMETER

CROSS-REFERENCE TO RELATED APPLICATIONS

[0001] This application is a divisional prior U.S. application Ser. No. 13/820,968, filed 4-Apr.-2013, which is a 371 of PCT/US11/577,747, filed 25-Oct.-2011, which claims the benefit of U.S. Provisional Application Nos. 61/406,251, filed 25-Oct.-2010, 61/406,255, filed 25-Oct.-2010, and 61/406,259, filed 25-Oct.-2010, which are each incorporated in its entirety by this reference.

TECHNICAL FIELD

[0002] This invention relates generally to the flow cytometer field, and more specifically to a new and useful systems and user interface in the flow cytometry field.

BACKGROUND

[0003] In a flow cytometer, light is directed onto a stream of sample fluid such that the light impinges and typically excites particles in the sample, causing the excited particles to emit light. The detection of the emitted light provides data that can be analyzed for characterizing the particles and the sample fluid, such as count, physical structure, chemical structure, and other useful information in applications such as for research and clinical purposes. The detection system is therefore a crucial component of a flow cytometer and is a factor in not only the quality (e.g., sensitivity, bandwidth) of the collected data, but also the overall structure and cost of the complete flow cytometer system. In conventional flow cytometers, the detection system includes photomultiplier tubes, or PMTs, which have relatively high sensitivity and high bandwidth, and produces data with relatively low noise. However, PMTs have several disadvantages, such as being relatively expensive and exhibiting temperature drift.

[0004] Furthermore, a typical flow cytometer detector has a limited collection range. In simple terms, the collection range of a typical flow cytometer is smaller than the signal range of the objects being analyzed with the flow cytometer. For this reason, the typical detector is supplied with a gain level and/or amplifier. Detectors typically collect data relative to an object's size (light scatter) or brightness (fluorescence); both types of data are often collected on each object detected in the sample. To collect signals from small or faint objects, the gain level is increased. With an increased gain level, however, the signals from large or bright objects are too intense to be collected. To collect signals from large or bright objects, the gain level is decreased. With a decreased gain level, however, the signals from small or faint objects are too weak to be collected. The setting of gain level and other parameters is complicated and difficult. The limitations of the user interface of typical flow cytometer systems have several disadvantages, including: (1) the expenditure of valuable user time spent on the gainsetting process to ensure it is set correctly; (2) the requirement of significantly more sample to determine the proper gain settings; (3) the potential loss of valuable data because at least a portion of input signals are outside of the user-set "active" dynamic collection range and are therefore not collected, and (4) the inability to observe and "undo" changes in user-set gain/scaling settings without running additional samples.

[0005] The use of detectors in flow cytometers is also complicated by complex optical systems. To use a conventional optical system, beam splitters and filters must be arranged in a very particular order to properly direct light of particular wavelengths to the appropriate detectors. Rearrangement of the optical system is required whenever a different wavelength detection configuration is required, such as experiments or tests using different fluorochromes. A user must skillfully perform this rearrangement, or the detector system will not function correctly. This limitation prevents the easy swapability of the filters and the easy modification of detection parameters. Further, the particular arrangement of the optical system decreases the reliability and the ruggedness of the flow cytometers, since alignment of the various optical components affects the operability of the detection system.

[0006] Thus, there is a need in the flow cytometry field to create new and useful systems and user interface. This invention provides such new and useful systems and user interface for collecting a data set in a flow cytometer.

BRIEF DESCRIPTION OF THE FIGURES

[0007] FIG. 1 is a schematic of the optical system of a preferred embodiment;

[0008] FIG. 2 is a schematic of a signal processing circuit for the detection system of a preferred embodiment;

[0009] FIG. 3 is a schematic of the detection system and virtual detector channels of preferred embodiments;

[0010] FIGS. 4A and 4B are flowcharts of a method for collecting a data set, enabled by the user interface of first and second preferred embodiments, respectively;

[0011] FIGS. 5A-5E are schematics of variations of a method for collecting a data set, enabled by the user interface of the first preferred embodiment; and

[0012] FIGS. 6A-6F are schematics of variations of a method for collecting a data set, enabled by the user interface of the second preferred embodiment.

DESCRIPTION OF THE PREFERRED EMBODIMENTS

[0013] The following description of preferred embodiments of the invention is not intended to limit the invention to these preferred embodiments, but rather to enable any person skilled in the art to make and use this invention.

[0014] In a preferred embodiment, systems for collecting a data set in a flow cytometer include optical and detection systems in a flow cytometer **102** having a flow channel with an interrogation zone and an illumination source that impinges the interrogation zone from a particular direction. As shown in FIG. 1, the optical system **100** preferably includes a lens subsystem including a collimating element that receives and collimates light from the interrogation zone, a light dispersion element that disperses the collimated light into a continuous wavelength spectrum of light, and a focusing lens that focuses light spectrum onto an array of adjacent detection points. The detection system **200** for a flow cytometer includes at least one semiconductor detector device that detects the focused light at a detection point and produces a signal corresponding to the detected light; a low noise amplifier circuit that amplifies the signal and is characterized by a high gain-bandwidth product; and a noise filter that reduces electronic noise from the amplified signal. The detection system **200** preferably includes multiple such

detectors in a detector array that collectively detects a full spectral range of input signals from the flow cytometer, in which each detector detects a subset spectral range of the full spectral range. The user interface **300** of a preferred embodiment enables a user to create a set of virtual detector channels by grouping detectors in the detector array, such that each virtual detector channel corresponds to a detector group and has a virtual detector channel range including the sum of subset spectral ranges of the detectors in the corresponding detector group. user interface of the preferred embodiments eliminate the discrete and complex system of detectors and filters used in conventional flow cytometers, thereby simplifying

[0015] The systems and the overall flow cytometer system and enabling a more compact, easier to use flow cytometer. The optical and detection systems also capture all usable light, thereby increasing the power and usability of the flow cytometer. Although the optical and detection systems and user interface are preferably integrated in a flow cytometer, the systems and user interface may alternatively be used in microscopy and/or any suitable apparatus or application for collecting and detecting light.

Optical System

[0016] As shown in FIG. 1, the optical system **100** preferably includes a lens subsystem including a collimating element **110** that receives and collimates light from the interrogation zone, a light dispersion element **120** that disperses the collimated light into a continuous wavelength spectrum of light, and a focusing lens **130** that focuses light spectrum onto an array of adjacent detection points **132**, such that a detector array may collect the focused light at the adjacent detection points **132**. The optical system **100** may further include a clean-up element that reduces spurious reflections and/or other undesired optical artifacts. In one embodiment, the optical system **100** receives and directs light towards a detection system that includes photodiode light detectors, but may alternatively receive and direct light toward a detection system that includes phototransistor light detectors, or any suitable detector.

[0017] The lens subsystem includes multiple lens and/or lens surfaces that function to manipulate the light from the illumination source into a form and/or to detection point locations where the light is detectable by a detector array. In a first stage of the lens subsystem, the collimating element **110** is preferably a collimating lens that aligns light from the interrogation zone, and more preferably an achromatic doublet lens, but may include any suitable combination of lenses or other collimating element. The collimating element **110** preferably perfectly or near-perfectly collimates the light into substantially parallel light rays, since the effects of imperfect collimation appear at subsequent stages of the lens subsystem, resulting in poorly-focused light at the detection points **132** and thereby degrading the detection of the light originating from the interrogation zone. As shown in FIG. 1, the collimating element **110** may be coupled to or otherwise optically paired with an aplanatic meniscus lens **112**. The pairing of the collimating element **110** and aplanatic meniscus lens **112** increases the effective numerical aperture of the first stage of the lens subsystem, thereby increasing the overall efficiency of light collection from the interrogation zone. The pairing also reduces the overall focal length of the first stage of the lens subsystem, without introducing additional undesirable spherical aberrations or other optical

aberrations, which enables the lens subsystem to be made more compact. In one exemplary embodiment, the collimating element **110** is an achromatic doublet lens with an effective focal length of approximately 30 millimeters and the aplanatic meniscus lens has an approximately 50 degree cone, such that the triplet lens combination of the collimating element and meniscus lens has an effective focal length of approximately 20 millimeters.

[0018] In a second stage of the lens subsystem, the light dispersion element **120** functions to disperse the collimated light from the collimating element **110** into a continuous wavelength spectrum of light. The light dispersion element **120** is preferably mounted in alignment with the output of the collimating element, such as in a frame, and may be permanently fixed or adjustable in position relative to the collimating element. As shown in FIG. 1, after passing through the light dispersion element **120**, the light is preferably redirected such that light rays of the same wavelength are parallel and light rays of different wavelengths are nonparallel at relative angles corresponding to their relative position in the wavelength spectrum. The light dispersion element **120** is preferably one of multiple variations. In a first variation, the light dispersion element **120** includes a diffraction grating that splits and diffracts light into a continuous spectrum of light rays towards the third stage of the lens subsystem. In a second variation, the light dispersion element **120** includes a dispersive prism that breaks up the collimated light into a continuous spectrum of light rays towards the third stage of the lens subsystem. In either variation, the dispersed light may follow a fold angle that is an acute angle, which may enable a more compact lens subsystem, or any suitable angle. In an exemplary embodiment, the light dispersion element **120** is a diffraction grating or prism with a grating of approximately 600 lines/millimeter, and the light dispersion element provides a fold angle of approximately 22 degrees. However, the light dispersion element may include any suitable structure with any suitable level of grating and/or fold angle.

[0019] In a third stage of the lens subsystem, the focusing lens **130** functions to focus the dispersed, continuous spectrum of light onto an array of adjacent detection points **132**. As shown in FIG. 1, the focusing lens **130** preferably gathers light of similar wavelength together such that all light focused on the array of detection points **132** is arranged in spectral order, but the dispersed light may be focused in any suitable manner. For example, beam splitters or additional stages in the lens subsystem may redirect selected spectral portions of the dispersed light to be arranged in any suitable spectral order, and/or focus selected spectral portions of the dispersed light onto detection points in any suitable location. Each detection point **132** may have a spot size of approximately 0.8 mm², or any suitable spot size. The array of detection points **132** is preferably a linear array. For instance, light corresponding to shorter wavelengths is preferably focused near a first end of the array of detection points **132**, and light corresponding to longer wavelengths is preferably focused near a second end of the array of detection points **132**, opposite the first end. However, the array of detection points may alternatively be an arcuate array (e.g., an open arc segment or enclosed circle or ellipse) or any suitable shape. In an exemplary embodiment, the focusing lens **130** has an effective focal length of approximately 160 millimeters and focuses the light to a linear array of detection points covering a dispersion distance of approximately

43 millimeters long. However, the focusing lens may be any suitable lens or lens surface, and may focus the light to any suitable arrangement of detection points.

[0020] In some embodiments, the optical system **100** may further include a cleanup element **140** that functions to reduce undesired optical artifacts or other aspects of the received light, such as crosstalk from spurious grating reflections. In one variation, the clean-up element **140** includes an optical filter. For example, the optical filter may be continuously variable, segmented, upper half, lower half, or any suitable type of filter. The filter may additionally and/or alternatively block one or more specific wavelengths from reaching the array of detector points **132**, such as the wavelength of a laser or lasers used to excite fluorescent particles in the interrogation zone. For example, such a blocking filter may include an optical notch filter, or a thin blocking bar that positioned at a particular detection point to absorb a particular wavelength or mask the detector from receiving light of the particular wavelength at the detection point, but the blocking filter may additionally and/or alternatively include any suitable filtering device. In another variation, the clean-up element **140** includes a slit grid, which may include any suitable slit grid known to one ordinarily skilled in the art or other suitable slit grid. The optical system may include any suitable number of clean-up elements **140**. One or more clean-up elements **140** may be positioned after the light dispersion stage of the lens subsystem and before or after the focusing stage of the lens subsystem, or in any suitable portion of the lens subsystem.

Detection System

[0021] The detection system **200** preferably includes one or more semiconductor light detector devices **212** in a detector array **210**. Each semiconductor detector device **212** detects the light at a respective detection point **132** (focused by the optical system **100** described above, or any suitable optical system) and produces an electrical signal corresponding to the detected light. As shown in FIG. 2, the detection system **200** preferably further includes signal processing circuitry including: a low noise amplifier circuit **220** that boosts or amplifies the signal from the detector device **212** and that is characterized by a high gain-bandwidth product; and a noise filter **230** that reduces electronic noise from the amplified signal.

[0022] The semiconductor detector device **212** functions to convert light into an electrical signal having characteristics that correlate with the nature of the light received from the interrogation zone. The wavelength sensitivity of the detector device **212** is preferably optimized to detect light between a full range of approximately 400 nanometers and 900 nanometers in wavelength, either a subset of the full range or over the entire full range. Furthermore, the wavelength sensitivity of the detector device **212** may be optimized to minimize detection of light in the infrared range, such as to avoid responding to sources of heat in the flow cytometer or other instrument in which the detection system is integrated. However, the detector devices may be optimized to detect any suitable range of wavelengths of light.

[0023] The angular sensitivity of the semiconductor detector device **212** may be optimized to receive light in a cone approximately corresponding to the focused cone of light emanating from the light focusing element of the optical system described above, while rejecting stray light passing outside of the cone, which may thereby minimize sensitivity

to light traveling off-axis within the optical system. Furthermore, the semiconductor detector device **212** may be shielded from any backlight or other ambient light sources to minimize background noise in the signal and improve the accuracy of the resulting collected data set from the flow cytometer (or other instrument). The semiconductor device **212** may additionally and/or alternatively be shielded from electromagnetic fields to minimize induced electronic noise. Furthermore, the semiconductor detector device **212** may include temperature compensation to minimize any temperature-induced changes in gain or linearity of the output signal.

[0024] The semiconductor detector device **212** may be any suitable semiconductor device or light detector device. In a preferred variation, the semiconductor detector device **212** is a photodiode, and more preferably a PIN photodiode, although the photodiode may be any suitable kind of diode. The photodiode preferably has a very low capacitance, such as approximately 20 pF or less. Since photodiodes with low capacitance are generally physically smaller and have a smaller light -sensitive region, the photodiode detector devices **212** are preferably placed at detection points **132** at which light is well-focused, such as by the optical system described above or any suitable optical system. The photodiode is preferably configured to output current that correlates with characteristics of the received light, but may alternatively be configured to output another suitable electrical characteristic, such as voltage.

[0025] In an alternative variation, the semiconductor detector device **212** is a phototransistor. Like the photodiode of the preferred variation, the phototransistor preferably has a very low capacitance and may have a small light -sensitive region such that the phototransistor requires placement at a detection point receiving well-focused light. The phototransistor may be configured to provide an output signal having a substantially linear gain response across the intended light power range of the input signals, such as by applying a suitable current bias to the base of the phototransistor, typically in the range of approximately 10 μ m to 1 mA. Furthermore, the output signal from the phototransistor may be converted from a current signal to a voltage signal, such as by feeding the output signal into the low-noise amplifier circuit, a load resistor or other suitable circuitry components. In some embodiments, the phototransistor may provide additional current gain that may be leveraged to increase gain of the signal and/or increase the signal-to-noise ratio. In particular, the use of a phototransistor may reduce the required resistance value of the gain resistor in the amplifier circuit, and consequently may reduce the thermal noise contribution to overall detector noise.

[0026] The low-noise amplifier circuit **220** functions to convert current input from the semiconductor detector device **212** to a voltage output and to amplify the signal from the semiconductor detector device **212**. The low-noise amplifier circuit **220** preferably includes a transimpedance amplifier, but may be any suitable kind of amplifier. The amplifier circuit preferably has a high open-loop gain-bandwidth product, such as approximately 1 GHz or more. The combination of a semiconductor detector device **212** having low capacitance and a transimpedance amplifier having a high gain-bandwidth product may enable the detection system **200** to have high gain (to amplify the signal from the detector) while maintaining high sensitivity and high bandwidth. The amplifier circuit **220** may further be

characterized with a high feedback resistance, low voltage noise and low current noise to reduce overall noise in the signal, and low input capacitance. In an exemplary embodiment, the transimpedance amplifier has a feedback resistance between 1-20 MOhms and preferably more than approximately 10 MOhms, voltage noise of less than approximately 5 nV/rtHz, current noise of less than approximately 5 fA/rtHz, and input capacitance less than approximately 6 pF. However, other embodiments of the amplifier circuit may have any suitable specifications. The amplifier circuit **220** may additionally and/or alternatively include any suitable electronic components that perform current-to-voltage conversion, or the detection system may include any suitable conversion circuitry (e.g., a passive current-to-voltage converter). The amplifier circuit may additionally and/or alternatively include digital signal processing.

[0027] The noise filter **230** functions to reduce electronic noise and/or photoninduced noise from the amplified signal, thereby increasing the signal-to-noise ratio. The noise filter **230** may include a low pass filter that quickly attenuates higher frequency noise above a predetermined cutoff point. The noise filter **230** may be implemented in hardware circuitry and/or digitally. In an exemplary embodiment, the low pass filter includes 120 dB or better attenuation, preferably with a 500 kHz or smaller transition band. The noise filter may additionally and/or alternatively include further signal processing techniques such as a smoothing algorithm.

[0028] As shown in FIG. 3, in a preferred embodiment, the detection system **200** preferably includes multiple semiconductor light detector devices **212** in a detector array **210** that collectively detects a full spectral range **250** of input signals from the flow cytometer, and in which each detector detects a subset spectral range **252** of the full spectral range. However, the semiconductor detector device **212** (e.g. photodiode or photo transistor) may alternatively be used in any suitable detection system. In an exemplary embodiment, the detection system detects a full spectral range **250** of input signals of approximately 400-900 nm, but may alternatively detect any suitable spectral range of light. Overall, the detection system **200** preferably is configured to have a bandwidth of at least approximately 400 kHz. However, in some embodiments, such as for use in flow cytometers having frequency modulated lasers at light sources in the interrogation zone, the detection system **200** may have a bandwidth of at least approximately 2 MHz. Alternatively, the detection system **200** may have any suitable bandwidth. Furthermore, the detector array **210** preferably detects a wide range of input signals that includes at least a 1:100,000 ratio, and more preferably at least a 1:1,000,000 ratio, between the faintest objects and the brightest objects. In a preferred embodiment, the detector array **210** is a linear array, such that a dispersed spectrum of light may be incident on the detector array, with light of shorter wavelengths detected near a first end of the linear array and light of longer wavelengths detected near a second end of the linear array opposite the first end. Alternatively, the detector array **210** may be an arcuate array (e.g., an open arc segment or enclosed circle or ellipse) or any suitable shape, preferably with adjacent and contiguous detector devices **212**.

[0029] Each detector device **212** in the detector array **210** preferably detects a respective portion or subset of the full spectral range **250** of the detection system **200**. In particular, the detector array **210** is preferably capable of detecting a continuous spectral range of light, and each detector device

212 may be configured to detect a subset spectral range based on its individual characteristics, and/or by its relative position in the detector array **210**. For example, the detector array **210** may include 50 detectors that each detects light of approximately 10 nm wavelength increments, from approximately 400-900 nm. The subset spectral ranges **252** of the detectors **212** may be of equal span (e.g. each detector detects, as a result of capability and/or respective position in the detector array, a subset range of 10 nm in the full spectral range), or may be of unequal span. In some embodiments, some or all of the detectors may detect light of overlapping spectra. For example, one detector may detect light of 510-530 nm, and an adjacent detector may detect light of 520-540 nm, such that the two adjacent detectors both detect light of 10 nm (520-530 nm). However, the detector array **210** may include any suitable number of detectors, and the detectors may detect any suitable wavelength ranges of light and/or overlapping wavelength ranges of light. The “spill-over” spectral overlap resulting from detection of a particular wavelength of light by multiple detectors may be automatically compensated by techniques known and used by one ordinarily skilled in the art, compensated by user-controlled techniques, and/or any suitable compensation methods. Although the detector array **210** preferably detects a continuous spectral range of light, the detector array may alternatively detect a discontinuous spectral range of light, or selected subsets of the full spectral range **250** of light.

[0030] As shown in FIG. 3, individual detectors may be grouped into virtual detector channels **240**, and/or actual detector channels, through which the data set from the flow cytometer or other instrument may be collected. Each virtual detector channel **240** corresponds to a detector group and has a virtual detector channel range **242** that includes the sum of the subset spectral ranges **252** of the detectors in the corresponding detector group. In other words, each virtual detector channel **240** includes the summed or combined input signals collectively detected by the individual detectors **212** in the detector group corresponding to the virtual detector channel **240**. Some or all of the virtual detector channels may include substantially equal virtual detector channel ranges or unequal virtual detector channel ranges. In particular, the detectors may be grouped in a first configuration corresponding to a first set of virtual detector channels **240** and in a second configuration corresponding to a second set of virtual detector channels **240**, where the first and second configurations are different. In other words, the grouping of signals from the detectors may be arranged and repeatedly rearranged in different groups between uses and applications, without requiring physical rearrangement of the components of the optical system and detection system.

User Interface

[0031] The user interface **300** for a flow cytometer is used to enable the grouping of detector signals to form virtual detector channels through which data is collected and organized. As shown in FIGS. 4A and 4B, in preferred embodiments, the user interface **300** provides a method for collecting a data set for a flow cytometer sample in a flow cytometer including the steps of: providing a detector array **8310** having a plurality of detectors that collectively detect a full spectral range of input signals from the flow cytometer, in which each detector detects a subset spectral range of the full spectral range; creating a set of virtual detector channels **8320** by grouping detectors in the detector array, and col-

lecting the full spectral range of input signals **8370** from the flow cytometer sample with the detector array. Each virtual detector channel corresponds to a detector group and has a virtual detector channel range that includes the sum of subset spectral ranges of the detectors in the corresponding detector group. The step of creating a set of virtual detector channels **8320**, which may include one or more of several variations of substeps, may be performed before and/or after the step of collecting input signals **8370**. The method may further include storing an initial data set based on the collected input signals **8380** and/or storing a configuration file of the configuration of the virtual detector channels **8390**. The user interface **300** enables extraction of data from a flow cytometer system having the optical and/or detection systems as described above, or any suitable instrument having an array of multiple detectors that each detects a portion of a full spectrum of light input signals. The user interface **300** enables a more comprehensive collection of data, and simplifies the process for setting up and configuring the detector system of the flow cytometer. However, the user interface and method for collecting a data set may alternatively be used in any suitable system requiring detection of a substantial spectrum of signals, such as microscopy.

[**0032**] The step of providing a detector array **8310** preferably includes providing a system that detects known subsets of an entire light (e.g., fluorescence) spectrum detected by the flow cytometer. The detector array is preferably similar to the detection system described above and shown in FIGS. **2** and **3**, but may alternatively be any suitable detection system. In particular, the detector array includes separate individual detector devices that detect a dispersed full spectrum of light such that adjacent detectors may detect portions or subset spectral ranges of the full spectrum of light, and more preferably such that adjacent detectors detect contiguous subset spectral ranges of the full spectrum of light. For example, within a group of adjacent detectors in the detector array, each detector may detect light in 10 nm increments, such that a first detector may detect light having a wavelength of approximately 491-500 nm, a second middle detector may detect light having a wavelength of approximately 501-510 nm, and a third detector may detect light having a wavelength of approximately 511-521 nm. The detector array may include photodiodes, phototransistors, or any suitable kind of light detector.

[**0033**] The step of creating a set of virtual detector channels **8320** functions to organize the signals collected by detectors in the detector array into designated data channels. As best shown in FIG. **3**, creating a set of virtual detector channels **8320** includes grouping detectors in the detector array **8322**. Each virtual detector channel **240** corresponds to a detector group and has a virtual detector channel range that includes the sum of the subset spectral ranges of the detectors in the corresponding detector group. In other words, each virtual detector channel **240** includes the summed or combined input signals collectively detected by the individual detectors in the detector group corresponding to the virtual detector channel. One or more detector groups may include detectors that are physically contiguous with each other (e.g., a “block” of detectors and/or detect contiguous subset spectral ranges of light, such that the corresponding virtual detector channel range collects light of a continuous spectral range (as in virtual detector channels **240a**, **240b**, and **240d**). Furthermore, one or more detector groups may

include detectors that are not physical contiguous with each (e.g., a “split block” of detectors) and/or detect not contiguous subset spectral ranges of light, such that the corresponding virtual detector channel range collects light of a discontinuous spectral range, as in virtual detector channel **240c**.

[**0034**] The step of collecting the full spectral range of input signals **8370** functions to gather raw data with the detector array. Collecting input signals **8370** may include collecting a full dynamic range of input signals that provides at least a 1:100,000 ratio, and more preferably at least a 1:1,000,000 ratio, between the faintest signals and the brightest signals from the flow cytometer sample. In a preferred embodiment, the data is collected in a raw, unmodified format without adjustment in gain level of the detectors, but may be collected in any suitable manner.

[**0035**] As shown in FIGS. **5** and **6**, the step of creating a set of virtual detector channels **S320** may be performed before and/or after the step of collecting input signals **S370**, such as before and/or after performing a sample run on the flow cytometer. In a first preferred embodiment, the step of creating a set of virtual detector channels **S320** is performed before collecting input signals. As shown in FIG. **5A**, in a first variation of this embodiment, the step of creating a set of virtual detector channels **S320** includes receiving a user selection of detector groups **S330** in which the user manually enters or indicates detector groups that make up the detector channels. For example, each detector may be labeled or numbered, and the user may specify that detectors “x” through “y” is a detector group. Since each detector is designated for detecting a particular subset spectrum range of input signals, the user selection of detector groups in turn creates virtual detector channels corresponding to the detector groups. For a flow cytometer system having multiple lasers, the user may further specify which laser is assigned to each detector, detector group, and/or virtual detector channel.

[**0036**] As shown in FIG. **5B**, in a second variation of the first preferred embodiment, the step of creating a set of virtual detector channels **S320** includes receiving a user selection of desired virtual detector channel ranges **S340** (or summed subset spectrum ranges of input signals) and grouping detectors into detector groups that correspond to the selection of virtual detector channel ranges **S342**, thereby forming a set of virtual detector channels that define the desired virtual detector channel ranges. For example, the user may specify a range of wavelengths, such as that between wavelength “a” and wavelength “b”, to assign to a particular virtual detector channel, and the user interface **300** may automatically correlate the range of wavelengths to specific individual detectors, by determining which individual detectors to group together to form a virtual detector channel that detects light between wavelengths “a” and “b”. In some embodiments, the user interface **300** may automatically exclude certain detectors from any detector group or virtual detector channel, such as in scenarios in which lasers in the optical system are active.

[**0037**] As shown in FIG. **5C**, in a third variation of the first preferred embodiment, the step of creating a set of virtual detector channels **8320** includes receiving one or more configuration parameters **8350** and optimally grouping detectors based on the configuration parameters **8352**. In an example of this variation, the step of creating a set of virtual detector channels includes receiving a user selection of a set of fluorochromes **8354** that tag a flow cytometer sample,

determining an optimal detector group for each fluorochrome based on at least one configuration parameter, and assigning the optimal detector group to each fluorochrome **8356** and thereby providing a virtual detector channel for each fluorochrome. Between different sample runs and/or sets of fluorochromes, the detector groups may be reassigned to provide different suitable virtual detector channels for different applications. Optimally grouping detectors **8352** may be dependent on one or more of several configuration parameters, including: minimizing spillover (overlapping detection of a fluorochrome between multiple detector channels), simplifying requirements for spillover compensation (typically algorithms to compensate for spillover in the data), instrument-specific calibration parameters, maximizing sensitivity of the detector channels (such as based on previous sample runs with a particular set of fluorochromes), any suitable configuration parameter based on the user-selected set of fluorochromes or instrumentation, or any suitable parameter. Optimally grouping detectors **8352** is preferably performed automatically by the user interface **300**, but may additionally and/or alternatively be performed manually by the user. The user interface and/or user may further determine the optical laser configuration in the flow cytometer for each fluorochrome.

[0038] As shown in FIG. 5csD, in a fourth variation of the first preferred embodiment, the step of creating a set of virtual detector channels **8320** includes receiving a configuration file **8360** that defines a predetermined group of detectors, arrangement of virtual detector channels, laser configuration, and/or any suitable settings for the flow cytometer system. In this variation, the step of grouping the detectors preferably incorporates the settings in the configuration file, and/or may include further modifications by the user or system. The configuration file may be directly provided by the user such as on portable media, selected from the user interface **300**, selected and downloaded from a network or server, provided in a machine-read sample label such as a bar code, or by any suitable means. The configuration file may be a saved configuration file from a previous sample run from the same or different flow cytometer system or other instrument, or may be a template configuration file. A configuration file defining the virtual detector channel settings and/or other instrument settings may help increase usability of the instrument, and help ensure consistency in analysis for similar experiments or tests. This consistency in analysis for similar experiments or tests may be particularly important in some applications, such as clinical applications.

[0039] In a second preferred embodiment, the step of creating a set of virtual detector channels **8320** is performed after collecting input signals **8370** (such as after a sample run with the flow cytometer system). As shown in FIGS. 6B and 6C, in first and second variations of this embodiment the step of creating a set of virtual detector channels **8320** include receiving a user selection of detector groups **8330** and receiving a user selection of desired virtual detector channel ranges **8340**, respectively, similar to the first and second variations of the first preferred embodiment.

[0040] As shown in FIG. 6D, in a third variation of this embodiment, the step of creating a set of virtual detector channels **8320** includes receiving one or more configuration parameters **8350** and optimally grouping detectors based on the configuration parameters **8352**, similar to the third variation of the first embodiment. Furthermore, the user interface **300** may additionally and/or alternatively optimize

the detector grouping based on bright and/or dim (or blank) peaks of a multi-intensity flow cytometer sample, which may be identified automatically by the user interface and/or manually by the user. This optimization may occur after a sample run, or after the user runs a set of experimental controls but before the actual sample run.

[0041] As shown in FIG. 6E, in a fourth variation of the second preferred embodiment, the step of creating a set of virtual detector channels **8320** includes receiving a configuration file **8360**, similar to the fourth variation of the first preferred embodiment.

[0042] Although the step of creating a set of virtual detector channels **8320** is preferably one of the above variations, in other embodiments the step of creating a set of virtual detector channels **8320** may be any suitable combination or permutation of the above variations and/or any suitable processes for forming virtual detector channels. For example, one or more of the variations of creating a set of virtual detector channels of the first embodiment may be implemented in some manner after the step of collecting input signals. Similarly, one or more of the variations of creating a set of virtual detector channels of the second embodiment may be implemented in some manner before the step of collecting input signals. Furthermore, through the user interface **300** or other means, the virtual detector channels may be created in multiple configurations for a single sample run (e.g. in a first configuration before collecting the input signals and in a second configuration, different from the first configuration, after collecting the input signals) and/or between different instances of sample runs (e.g. in a first configuration for one sample run and in a second configuration, different from the first configuration, for another sample run).

[0043] As shown in FIGS. 4-6, the user interface **300** may further enable storing the collected input signals from most, if not all, available detectors. Storing the collected input signals may include storing an input signal as the signal individually collected by each separate detector in the detector array **S380** (e.g., FIG. 6A), and/or may include storing the input signal as the signal collected by each separate detector group through the virtual detector channel **S380'** (e.g., FIG. 6F). The data set of collected input signals may be saved to a local memory, portable media, server, network or any suitable memory. The stored data set preferably includes the full spectral range of the detectors, but may include any suitable portion of the spectral range or collected signals. The stored data set may also be useful for analysis and other manipulations, as described in U.S. Pat. No. 7,996,188 entitled "User interface for a flow cytometer", which is incorporated in its entirety by this reference.

[0044] As shown in FIGS. 4-6, the user interface **300** may further enable storing a configuration file **S390** that defines a predetermined grouping of detectors, arrangement of virtual detector channels, laser configuration, and/or any suitable settings for the flow cytometer system. The step of storing the configuration file **S390** may include saving to a local memory or portable media, saving to a server or network, or any suitable saving step. The stored configuration file may be used for future sample runs, as described above, and/or as a template. Multiple configuration files may be stored. For example, a first configuration file may be stored immediately after a sample run to save an initial configuration of detector groups and virtual detector channels, and a second configuration file may be stored after

optimizing the grouping based on characteristics of the collected data. In this example, the first and second configuration files may be compared or analyzed for future reference. The user interface 300 may additionally and/or alternatively include exporting the configuration file to a different medium, such as a printout.

[0045] As a person skilled in the art will recognize from the previous detailed description and from the figures and claims, modifications and changes can be made to the preferred embodiments of the invention without departing from the scope of this invention defined in the following claims.

1-9. (canceled)

10. A system comprising:

a light source for irradiating a sample in a flow stream;
a light dispersion component configured to disperse a spectrum of light from the sample; and

a light detection component comprising:

a plurality of detectors; and

a plurality of detector channels, wherein each detector channel corresponds to one or more detectors.

11. The system according to claim 10, wherein each detector channel corresponds to a predetermined group of detectors.

12. The system according to claim 10, wherein each detector channel corresponds to a user-assigned group of detectors.

13. The system according to claim 12, wherein the detectors of each detector channel are assigned based on a fluorophore present in the sample.

14. The system according to claim 12, wherein the detectors of each detector channel are assigned to minimize spillover between detector channels.

15. The system according to claim 10, wherein the dispersion component comprises a dispersive prism.

16. The system according to claim 10, wherein the dispersion component comprises a diffraction grating.

17. The system according to claim 10, wherein the light detection component comprises an array of detectors.

18. The system according to claim 17, wherein the light detection component comprises a linear array of detectors.

19. The system according to claim 17, wherein the light detection component comprises an array of photodiodes.

20. The system according to claim 17, wherein the light detection component comprises an array of phototransistors.

21. The system according to claim 17, wherein the array of detectors is configured to detect light from 400 nm to 900 nm.

22. The system according to claim 21, wherein the array of detectors comprises 50 or more detectors that each detect light in 10 nm increments.

23. The system according to claim 22, wherein the spectral range of each detector in the array overlaps with adjacent detectors.

24. The system according to claim 10, wherein light detection component further comprises an amplifier circuit comprising a plurality of amplifiers configured to convert current from each detector channel into a voltage output signal.

25. The system according to claim 10, wherein light detection component further comprises an optical adjustment component.

26. The system according to claim 25, wherein the optical adjustment component comprises a lens that propagates light from the sample.

27. The system according to claim 26, wherein the optical adjustment component comprises a focusing lens.

28. The system according to claim 10, wherein the light source comprises a laser.

29. A method comprising:

analyzing a sample with a system comprising:

a light source for irradiating the sample in a flow stream;
a light dispersion component configured to disperse a spectrum of light from the sample; and

a light detection component comprising:

a plurality of detectors; and

a plurality of detector channels, wherein each detector channel corresponds to one or more detectors.

* * * * *